

# Instructions for Use

## SHIKARI® (S-ATA)

# Semi-Quantitative Free/Total Antibodies to Adalimumab ELISA

Enzyme immunoassay for the semi-quantitative free and total antibodies to adalimumab in serum and plasma samples

<b>REF</b>	ADA-QNFT-HUM
	96 tests
	Shipment 10-30°C, Store 2-8°C
	MATRIKS BIOTECHNOLOGY CO., LTD. Gazi Universitesi Teknoplaza CZ-26/27 06830 Golbasi Ankara / TURKEY Tel +90 312 485 42 94 <a href="mailto:info@matriksbiotek.com">info@matriksbiotek.com</a>

### 1. Intended Use

SHIKARI® Semi-Quantitative Free/Total Antibodies to Adalimumab ELISA has been especially developed for the semi-quantitative analysis of free and total antibodies to adalimumab in serum and plasma samples. SHIKARI® Semi-Quantitative Free/Total Antibodies to Adalimumab ELISA is optimized with Humira®.

### 2. General Information

Adalimumab binds with specificity to tumour necrosis factor-alpha (TNF-alpha) and inhibits its interaction with the p55 and p75 cell surface TNF receptors. Adalimumab also lyses surface tumour necrosis factor expressing cells in vitro when in the presence of complement. Adalimumab does not bind or inactivate lymphotoxin (Tumour necrosis factor-beta). TNF is a naturally occurring cytokine that plays a role in normal inflammatory and immune responses. Increased levels of TNF are found in the joint synovial fluid of rheumatoid arthritis, psoriatic arthritis, and ankylosing spondylitis patients, and play an imperative role in pathologic inflammation and the joint destruction that are major complications of these diseases. Increased levels of TNF are also measured in psoriasis plaques. In plaque psoriasis, treatment with adalimumab may decrease the epidermal thickness and inflammatory cell infiltration. The relationship between this pharmacodynamics and the mechanism(s) by which adalimumab achieves its clinical effects is not known. Additionally, adalimumab alters biological responses that are induced/regulated by TNF, including changes in the levels of adhesion molecules responsible for leukocyte migration during inflammation (ELAM-1, VCAM-1, and ICAM-1 with an IC50 of 1-2 X 10-10M).

Therapeutic drug monitoring (TDM) is the clinical practice of measuring specific drugs at designated intervals to maintain a constant concentration in a patient's bloodstream, thereby optimizing individual dosage regimens. The indications for drug monitoring include efficacy, compliance, drug-drug interactions, toxicity avoidance, and therapy cessation monitoring. Additionally, TDM can help to identify problems with medication compliance among noncompliant patient cases.

Biologic medicinal products (biologics) have transformed treatment landscapes worldwide for patients with haematological or solid malignancies with the 21st century. Today, as data exclusivity periods of first wave biologics approach expiration/have expired, several biosimilar products (i.e., biologics that are considered to be similar in terms of quality, safety and efficacy to an approved 'reference' biologic) are being developed or have already been approved for human use.

Like all biologics, biosimilars are structurally complex proteins that are typically manufactured using genetically engineered animal, bacterial or plant cell culture systems. As a consequence of this molecular complexity and the proprietary nature of the manufacturing process, which will inevitably result in the use of different host cell lines and expression systems as well as related differences in manufacturing conditions, it is not possible to manufacture exact copies of a reference biologic.

When administered to patients, all therapeutic proteins have the potential to induce an unwanted immune response (i.e., to stimulate the formation of antidrug antibodies [ADAs]). The impact of immune responses can range from no apparent effect to changes in pharmacokinetics, loss of effect and serious adverse events. Furthermore, the immunogenicity profile of a biologic can be significantly altered by even small differences in its manufacturing process that are accompanied by a change in product attributes, as well as differences in dosing schedules, administration routes or patient populations.

SHIKARI® ELISA kits can be used for drug level and anti-drug antibodies measurements. SHIKARI® adalimumab ELISA products:

Brand	Description	Product Code
SHIKARI® (Q-ADA)	Adalimumab Free drug	ADA-FD-HUM
SHIKARI® (S-ATA)	Adalimumab Antibody screening - Qualitative	ADA-QLS-HUM
SHIKARI® (S-ATA)	Adalimumab Antibody screening - Quantitative	ADA-QNS-HUM
SHIKARI® (S-ATA)	Adalimumab Antibody screening - Free/Total semi-quantitative	ADA-QNFT-HUM

Check the web page for the whole product list [www.matriksbiotek.com](http://www.matriksbiotek.com)

### 3. Test Principle

Shikari® Semi-Quantitative Free/Total Antibodies to Adalimumab ELISA is a sandwich assay for the determination of total and free antibodies against adalimumab in serum and plasma samples. During the first incubation period, the separation of adalimumab specific antibody-adalimumab immune complex is provided by adding dissociation buffer. After transferring dissociation mix to the plate, adalimumab antibodies are separated from adalimumab in patient serum/plasma samples and they are captured by the drug adalimumab coated on the wall of the microtiter wells and horse radish peroxidase (HRP) conjugated probe. After washing away the unbound components from samples, the bound enzymatic activity is detected by addition of tetramethylbenzidine (TMB) chromogen substrate. Finally, the reaction is terminated with an acidic stop solution. The intensity of the reaction colour is directly proportional to the concentration of adalimumab antibodies in sample.

Shikari® Semi-Quantitative Free/Total Antibodies to Adalimumab ELISA kit can be also used as a semi-quantitative test for free anti-drug antibodies determination without dissociation and neutralization steps. Peroxidase labelled specific conjugate and diluted serum/plasma samples are transferred simultaneously to the adalimumab-coated plate and antibodies to adalimumab in patient serum/plasma samples are captured by the drug adalimumab coated on the wall of the microtiter wells and HRP conjugated probe. After washing away the unbound components from samples, the bound enzymatic activity is detected by addition of TMB chromogen substrate. Finally, the reaction is terminated with an acidic stop solution. The intensity of the reaction colour is directly proportional to the concentration of adalimumab antibodies in sample.

### 4. Warnings and Precautions

- For professional use only.
- In case of severe damage of the kit package please contact Matriks Biotek® or your supplier in written form, latest one week after receiving the kit. Do not use damaged components in test runs but keep safe for complaint related issues.
- Obey lot number and expiry date. Do not mix reagents of different lots. Do not use expired reagents.
- Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood. For further information (clinical background, test performance, automation protocols, alternative applications, literature, etc.) please refer to the local distributor.
- Follow good laboratory practice and safety guidelines. Wear lab coats, disposable latex gloves and protective glasses where necessary.
- All reagents of this kit containing human serum or plasma (standards etc.) have been tested and were found negative for HIV I/II, HBSAg and Anti-HCV. However, a presence of these or other infectious agents cannot be excluded absolutely and therefore reagents should be treated as potential biohazards in use and for disposal.
- Reagents of this kit containing hazardous material may cause eye and skin irritations. See "Materials supplied", MSDS and labels for details.
- Chemicals and prepared or used reagents must be treated as hazardous waste according to the national biohazard safety guidelines or regulations.

### 5. Storage and Stability

The kit is shipped at ambient temperature (10-30°C) and should be stored at 2-8°C for long term storage. Keep away from heat or direct sunlight. The strips of microtiter plate are stable up to the expiry date of the kit in the broken, but tightly closed bag when stored at 2-8°C.

### 6. Specimen (Collection and Storage)

Serum, Plasma (EDTA, Heparin)

The usual precautions for venipuncture should be observed. Do not use grossly haemolytic, icteric or lipemic specimens. Samples appearing turbid should be centrifuged before testing to remove any particulate material. Avoid repeated freeze-thaw cycles for serum/plasma samples.

Samples should be diluted with the dilution rate given in the "Pre-test setup instructions" before the test.

Drug infusions may camouflages/mask the presence of antibody to drugs in serum/plasma samples. Therefore, blood sampling time is critical for detection of antibodies. It is recommended to take the blood sample just before the scheduled dose (trough specimen).

Storage	2-8°C	-20°C
Stability (serum/plasma)	2 days	6 months

### 7. Materials Supplied

Material	Quantity	Description
Microtiter Plate	1 x 12 x 8	Microtiter plate Break apart strips. Microtiter plate with 12 rows each of 8 wells coated with adalimumab.
Controls	1,0 mL (negative) 0,5 mL (positive)	Control negative and positive Ready to use. Contains human serum and stabilizer, <0,1% NaN <sub>3</sub> .
Assay Buffer	1 x 50 mL	Assay buffer Ready to use. Blue coloured. Contains proteins, <0,1% NaN <sub>3</sub> .
Immune Complex Control	1 x 1,5 mL	Immune complex control Ready to use. Contains anti-adalimumab/adalimumab immune complex, human serum and stabiliser. <0,1% NaN <sub>3</sub> .
Dissociation Buffer	1 x 25 mL	Dissociation buffer Ready to use. Contains diluted acid.
Neutralisation Buffer	1 x 5 mL	Neutralisation buffer Ready to use.
Conjugate	1 x 12 mL	Horse radish peroxidase conjugated probe Ready to use. Red coloured. Contains HRP conjugated probe, stabilizer and preservatives.
Substrate	1 x 12 mL	TMB substrate solution Ready to use. Contains 3,3',5,5'-Tetramethylbenzidine (TMB).
Stop Buffer	1 x 12 mL	TMB stop solution Ready to use. 1N HCl.
Wash Buffer	1 x 50 mL	Wash buffer (20x) Prepared concentrated (20x) and should be diluted with the dilution rate given in the "Pre-test setup instructions" before the test. Contains buffer with tween 20.
Foil	2 x 1	Adhesive Foil For covering microtiter plate during incubation

### 8. Materials Required but Not Supplied

- Micropipettes and tips
- Calibrated measures
- Tubes for sample dilution
- Wash bottle, automated or semi-automated microtiter plate washing system
- Microtiter plate reader capable of measuring optical density with a photometer at OD 450nm with reference wavelength 650 nm (450/650 nm)
- Distilled or deionised water, paper towels, pipette tips and timer

### 9. Procedure Notes

- Any improper handling of samples or modification of the test procedure may influence the results. The indicated pipetting volumes, incubation times, temperatures and pre-treatment steps must be performed strictly according to the instructions. Use calibrated pipettes and devices only.
- Once the test has been started, all steps should be completed without interruption. Make sure that required reagents, materials and devices are prepared ready at the appropriate time. Allow all reagents and specimens to reach room temperature (18-25°C) and gently swirl each vial of liquid reagent and sample before use. Mix reagents without foaming.
- Avoid contamination of reagents, pipettes and wells/tubes. Use new disposable plastic pipette tips for each reagent, standard or specimen. Do not interchange caps. Always cap not used vials. Do not reuse wells/tubes or reagents.
- Use a pipetting scheme to verify an appropriate plate layout.
- Incubation time affects results. All wells should be handled in the same order and time sequences. It is recommended to use an eight-channel micropipette for pipetting of solutions in all wells.
- Microplate washing is important. Improperly washed wells will give erroneous results. It is recommended to use a multichannel pipette or an automatic microplate washing system. Do not allow the wells to dry between incubations. Do not scratch coated wells during rinsing and aspiration. Rinse and fill all reagents with care. While rinsing, check that all wells are filled precisely with wash buffer, and that there are no residues in the wells.
- Humidity affects the coated wells/tubes. Do not open the pouch until it reaches room temperature. Unused wells/tubes should be returned immediately to the resealed pouch including the desiccant.

### 10. Pre-test Setup Instructions

#### – Preparation of components

<b>Component</b>	Wash buffer (must be prepared before starting assay procedure)
<b>Dilute</b>	10 mL (e.g.)
<b>With</b>	Up to 200 mL
<b>Diluent</b>	Distilled water
<b>Dilution Ratio</b>	1/20
<b>Remarks</b>	Warm up 37°C to dissolve crystals. Mix vigorously.
<b>Storage</b>	2-8°C
<b>Stability</b>	2 weeks

#### – Dilution of samples and controls for free antibodies

<b>Sample</b>	Serum/Plasma	Controls
<b>Diluent</b>	Assay buffer	Assay buffer
<b>Dilution Ratio</b>	1/5	1/5
<b>Remarks</b>	1/5 dilution 40 µL sample + 160 µL assay buffer	1/5 dilution 40 µL control + 160 µL assay buffer

#### – Dilution of immune complex control for total antibodies

<b>Sample</b>	Immune complex control
<b>Diluent</b>	Assay buffer
<b>Dilution Ratio</b>	1/5
<b>Remarks</b>	1/5 dilution 40 µL control + 160 µL assay buffer

### 11. Test Procedure

Total Antibodies to Adalimumab	Free Antibodies to Adalimumab
Total assay time: 95 minutes	Total assay time: 80 minutes
Pipette 40 µL of each "Negative control", "Positive control" and "Immune complex control" and samples into the respective tubes	
Add 160 µL "Dissociation buffer" to tubes and incubate the plate 15 minutes at room temperature (18-25°C)	
Pipette 65 µL "Peroxidase conjugate" and 35 µL "Neutralisation buffer" into each of the wells to microtiter plate and transfer 100 µL dissociation mix into each of the respective wells of microtiter plate	Pipette 65 µL "Peroxidase conjugate" into each of the wells to microtiter plate and transfer 135 µL of diluted "Negative control", "Positive control" and samples into each of the respective wells of microtiter plate
Wells A1: Negative control* B1: Negative control* C1: Positive control D1: Immune complex control (after acid dissociation) E1: Immune complex control (before acid dissociation, diluted with assay buffer) F1 and on: Samples	Wells A1: Negative control* B1: Negative control* C1: Positive control D1 and on: Samples
*It is advised to run more than one "Negative control" samples and. Negative control studies can be duplicated or triplicated in order to take the mean value	*It is advised to run more than one "Negative control" samples and. Negative control studies can be duplicated or triplicated in order to take the mean value
Common Steps	
Cover the plate with adhesive foil	
Briefly mix contents by gently shaking the plate	
Incubate 60 minutes at room temperature (18-25°C)	
Remove adhesive foil	
Discard incubation solution	
Wash plate three times each with 300 µL "Wash Buffer"	
Remove excess solution by tapping the inverted plate on a paper towel	
Pipette 100 µL "Substrate" into each well	
Incubate 20 minutes without adhesive foil at room temperature (18-25°C) in the dark	
Stop the substrate reaction by adding 100 µL "Stop Solution" into each well	
Briefly mix contents by gently shaking the plate	
Colour changes from blue to yellow	
Measure optical density with a photometer at OD 450nm with reference wavelength 650 nm (450/650 nm) within 30 minutes after pipetting the "Stop Solution"	

### 12. Quality Control

The test results are only valid if the test has been performed following the instructions. Moreover, the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable standards/laws. For the run to be valid, the OD 450/650 nm of positive control should be >1,500 and the OD 450/650 nm of each negative control should be <0,150. In case of any deviation the following technical issues (but not limited to) should be reviewed: Expiration dates of reagents, storage conditions, pipettes, devices, incubation conditions, washing methods, etc.

### 13. Calculation and Interpretation of Results

- The results are expressed in arbitrary units (AU/mL, antibody unit/mL)
- The results are evaluated by a cut-off calculation is as follows

Cut-off value:  $2 \times \text{Negative control OD } 450/650 \text{ nm} = 10 \text{ antibody units (AU/mL)}$

Range	Interpretation
$\geq 10 \text{ AU/mL}$	Positive
$< 10 \text{ AU/mL}$	Negative

Note: The cut-off information provided with this kit can only be considered as a recommendation. Cut-off values must be calculated/set or verified according to scientific standards by the users/laboratories.

e.g.

Patients sample OD 450/650 nm: 0,600

The mean OD 450/650 nm of negative controls: 0,075

Cut-off value =  $2 \times 0,075 = 0,150 = 10 \text{ AU/mL}$

Result of the patient sample:  $0,600 / 0,150 = 4 \times 10 \text{ AU/mL} = 40 \text{ AU/mL}$ , Positive

- Immune complex control interpretation

Acid dissociation (AD) control is done with immune complex control (ICC) results.

$$\frac{\text{After AD ICC OD } 450/650 \text{ nm} - \text{Before AD ICC OD } 450/650 \text{ nm}}{\text{Before AD ICC OD } 450/650 \text{ nm}} \times 100 > \%20$$

e.g.

Before AD ICC OD 450/650 nm: 0,286

After AD ICC OD 450/650 nm: 0,050

$(0,286 - 0,050) / 0,050 \times 100 = \%472$

### 14. Analytical Performance

- Specificity: There is no cross reaction with native serum immunoglobulin
- Precision: Intra-assay and inter-assay CVs <30%
- Cut-off: Cut-off values must be calculated/set or verified according to scientific standards by the users/laboratories.

The "Quality control certificate" contains lot specific analytical performance data and is supplied separately with each kit. If some further analytical performance data is needed, please refer to the local distributor.

### 15. Automation

SHIKARI® Semi-Quantitative Free/Total Antibodies to Adalimumab ELISA is also suitable to run on automated ELISA processors.

### 16. Symbols and Cautions

	Manufacturer		Temperature limitation
	Production date		See instruction for use
	Expiry date		Caution
	Lot number		In vitro diagnostic medical device
	Catalog number		Control
	Do not use if package is damaged		Negative control
	Keep away from sunlight		Positive control
	Keep dry		Number of tests

According to ISO 15223

**Cautions:** The performance of the kit can be achieved by fully complying with the instructions. Modifications on the test procedure can affect the results and these kinds of changes will not be charged as regular complaints. This product is for professional use only and must be used for "Intended use" that is given in the instructions for use. The results themselves should not be the only reason for any therapeutically consequences. They must be correlated to other clinical observations. Cut-off, reference ranges, etc. must be calculated/set according to scientific standards by the users/laboratories. Information in the instructions about cut-off, etc. performance characteristics, can only be considered as a recommendation and does not give any responsibility to the manufacturer.

**Limitations of liability:** The manufacturer's liability is limited to the purchase price of the product in all circumstances. The manufacturer cannot be held responsible for damage to the patient, lost profit, lost sales, damage to property or any other incidental or consequential loss.

**Technical support and complaints:** Technical support can be given upon request. If there is a problem with the product, complaints must be sent written to [info@matriksbiotek.com](mailto:info@matriksbiotek.com) with the technical data (if available) like standard curve, control results, etc. After the necessary examination, written reply will be given.

### 17. References

- Mease PJ. Adalimumab in the treatment of arthritis. Ther Clin Risk Manag 2007 Mar;3(1):133-48.
- Scheinfeld N. Adalimumab (HUMIRA): A review. J Drugs Dermatol 2003 Aug;2(4):375-7.
- Humira FDA Label

### 18. Revision summary

Revision no	Release date	Explanation
IFU-033-01	03.02.2020	New documentation

