



innovation for health & wellness

“trace & catch”

Instructions or Use

Antibody o Rituximab

(Rituxan[®], Mabthera[®]) ELISA

SHIKARI[®] S-ATR

Enzyme immunoassay for the qualitative determination of antibodies o rituximab in serum and plasma

REF TR-ARTXv1  12 x 8    2-8°C

Revision # 1.1 December 2016



Matriks Biotek[®] Laboratories
www.matriksbiotek.com

Contents	Page
Intended Use	3
Summary and Explanation	3
Test Principle	6
Warnings and Precaution	6
Storage and Stability.....	7
Specimen Collection and Storage	8
Materials Supplied.....	8
Materials Required but not Supplied	9
Procedure Notes.....	9
Preparation of Component	10
Test Procedure.....	10
Interpretation of Results	11
References	11

	SHIKARI S-ATR
	Rituximab (Rituxan®, Mabthera®) antibodies qualitative analyses
Required Volume (µl)	10
Total Time (min)	140
Sample	Serum, plasma
Sample Number	96
Detection Limit (ng/mL)	+/-
Spike Recovery (%)	-
Shelf Life (year)	1

Intended Use

The *Matriks Biotek*® Antibody to Rituximab (Rituxan®, Mabthera®)* Enzyme-Linked-Immuno-Sorbent-Assay (ELISA) Kit is intended for the qualitative determination of antibodies to rituximab (Rituxan®, Mabthera®) in serum and plasma. It is for professional use only.

* Rituxan® is a registered trademark of Biogen Idec, Inc., Mabthera® is a registered trademark of Roche, Inc.,

Summary and Explanation

Rituximab is a genetically engineered chimeric murine/human monoclonal antibody directed against the CD20 antigen found on the surface of normal and malignant B lymphocytes. The antibody is a glycosylated IgG1 kappa immunoglobulin containing murine light- and heavy-chain variable region sequences (Fab domain) and human constant region sequences (Fc domain). Rituximab is composed of 1,328 amino acids and has an approximate molecular weight of 144 kD. Rituximab has a high binding affinity for the CD20 antigen of 5.2 to 11.0 nM.

Rituximab binds specifically to the antigen CD20, a transmembrane molecule located on pre-B and mature B lymphocytes. The antigen is expressed on > 95% of all B-cell non-Hodgkin's lymphomas (NHL). CD20 (human B lymphocyte-restricted differentiation antigen, Bp35) is a hydrophobic transmembrane protein with a molecular weight of approximately 35 kD. This non-glycosylated phosphoprotein is found on both normal and malignant B cells, but not on haematopoietic stem cells, pro-B cells, normal plasma cells or other normal tissues. CD20 regulates (an) early step(s) in the activation process for cell cycle initiation and differentiation, and possibly functions as a calcium ion channel. CD20 does not internalise upon antibody binding and is not shed from the cell surface. This antigen does not circulate in the plasma. Thus, free antigen does not compete for rituximab binding.

The Fab domain of rituximab binds to the CD20 antigen on B-lymphocytes and the Fc domain recruits immune effector functions to mediate B-cell lysis. Possible mechanisms of cell lysis include complement-dependent cytotoxicity (CDC) and antibody-dependent cellular cytotoxicity (ADCC). The antibody also induces apoptosis in the DHL-4 human B-cell lymphoma line and in vitro studies have demonstrated that rituximab sensitises drug-resistant human B-cell lymphoma lines to the cytotoxic effects of some chemotherapeutic agents. In human tissue, the expression of the CD20 antigen is highly restricted; rituximab binding to CD20 was found only on lymphoid cells in the thymus,

mg dose and ranged from 355 to 404 µg/mL for the 2 x 1000 mg dose. Mean terminal elimination half-life ranged from 15 to 16.5 days for the 2 x 500 mg dose group and 17 to 21 days for the 2 x 1000 mg dose group. Mean C_{max} was 16 to 19% higher following second infusion compared to the first infusion for both doses. Upon re-treatment with a second course the pharmacokinetics of rituximab were again assessed following two IV doses of 500 mg and 1000 mg. Mean C_{max} for serum rituximab following first infusion was 170 to 175 µg/mL for 2 x 500 mg dose and 317 to 370 µg/mL for 2 x 1000 mg dose. C_{max} following second infusion, was 207 µg/mL for the 2 x 500 mg dose and ranged from 377 to 386 µg/mL for the 2 x 1000 mg dose. Mean terminal elimination half-life after the second infusion, following the second course, was 19 days for 2 x 500 mg dose and ranged from 21 to 22 days for the 2 x 1000 mg dose. PK parameters for rituximab were comparable over the two treatment courses.

Wegener's Granulomatosis(WG) and Microscopic Polyangiitis(MPA): Based on the population pharmacokinetic analysis of data in 97 WG and MPA patients who received 375 mg/m² rituximab once weekly by intravenous infusion for four weeks, the estimated median terminal elimination half-life was 23 days (range, 9 to 49 days). Rituximab mean clearance and volume of distribution were 0.312 L/day (range, 0.115 to 0.728 L/day) and 4.50 L (range, 2.21 to 7.52 L) respectively. Male patients and patients with higher BSA or positive human anti-chimeric antibody (HACA) levels have higher clearance. However, further dose adjustment based on gender or HACA status is not necessary.

Immunogenicity

As with all therapeutic proteins, there is a potential for immunogenicity. The observed incidence of antibody (including neutralizing antibody) positivity in an assay is highly dependent on several factors including assay sensitivity and specificity, assay methodology, sample handling, timing of sample collection, concomitant medications, and underlying disease. It was reported that, using an ELISA assay, anti-human anti-chimeric antibody (HACA) was detected in 4 of 356 (1.1%) patients with low-grade or follicular NHL receiving single-agent rituximab. Three of the four patients had an objective clinical response. A total of 273/2578 (11%) patients with RA tested positive for HACA at any time after receiving rituximab. HACA positivity was not associated with increased infusion reactions or other adverse reactions. Upon further treatment, the proportions of patients with infusion reactions were similar between HACA positive and negative patients, and most reactions were reported to be mild to moderate. Four HACA positive patients had serious infusion reactions, and the temporal relationship between HACA positivity and infusion reaction was variable. A total of 23/99 (23%) rituximab-treated patients with WG and MPA tested positive for HACA by 18 months. The clinical relevance of HACA formation in rituximab-treated patients is unclear.

In rituximab-treated cynomolgus monkeys B lymphocyte numbers were reduced by 99% or more in comparison with pre-test values in the peripheral blood of all monkeys, approximately 24 hours after the first dose. Two weeks after the last dose, B lymphocyte numbers were still reduced by more than 99% in 3/6 monkeys dosed for four weeks and in 4/6 monkeys dosed for eight weeks, and B lymphocyte numbers were also depleted in the mandibular lymph nodes and femoral bone marrow. However, a partial recovery of B lymphocyte numbers in the peripheral blood of some monkeys in both dose groups was correlated with the development of antibodies against rituximab.

The use of rituximab (Rituxan[®], Mabthera[®]) was associated to the development of anti-rituximab antibodies, even some might be neutralizing, in various percentages of patients during therapy with the drug. The Matriks Biotek[®] Rituximab-ELISA and Antibody to rituximab ELISA Kits can be efficiently used, for monitoring serum through levels and the presence of anti-rituximab antibodies respectively, during therapy and offers the scientist a tool for decision on possible preventive measures.

Test Principle

The *Matriks Biotek[®] Antibody to Rituximab (Rituxan[®], Mabthera[®]) ELISA* is a double-antigen sandwich assay for the determination of antibodies against rituximab in serum and plasma samples. During the first incubation period, the drug rituximab coated on the wall of the microtiter wells captures the antibodies to rituximab in patient serum and plasma samples. After washing away the unbound components from samples, a Peroxidase-rituximab conjugate is added to each well and then incubated. After a second washing step, the bound enzymatic activity is detected by addition of tetramethylbenzidine (TMB) chromogen-substrate. Finally, the reaction is terminated with an acidic stop solution. The intensity of the reaction color is directly proportional to the concentration of antibodies of rituximab in sample.

Warnings and Precautions

1. For professional use only.
2. Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood. For further information (clinical background, test performance, automation protocols, alternative applications, literature, etc.) please refer to the local distributor.

3. In case of severe damage of the kit package please contact Matriks Biotek or your supplier in written form, latest one week after receiving the kit. Do not use damaged components in test runs, but keep safe for complaint related issues.
4. Obey lot number and expiry date. Do not mix reagents of different lots. Do not use expired reagents.
5. Follow good laboratory practice and safety guidelines. Wear lab coats, disposable latex gloves and protective glasses where necessary.
6. Reagents of this kit containing hazardous material may cause eye and skin irritations. See MATERIALS SUPPLIED and labels for details.
7. Chemicals and prepared or used reagents have to be treated as hazardous waste according the national biohazard safety guidelines or regulations
8. Avoid contact with Stop solution. It may cause skin irritations and burns
9. Some reagents contain sodium azide (NaN_3) as preservatives. In case of contact with eyes or skin, flush immediately with water. NaN_3 may react with lead and copper plumbing to form explosive metal azides. When disposing reagents, flush with large volume of water to avoid azide build-up.
10. All reagents of this test kit containing human serum or plasma have been tested and were found negative for HIV I/II, HBsAg and HCV by FDA approved procedures. However, a presence of these or other infectious agents cannot be excluded absolutely and therefore reagents should be treated as potential biohazards in use and for disposal.

Storage and Stability

The kit is shipped at ambient temperature and should be stored at 2-8°C. Keep away from heat or direct sun light. The storage and stability of specimen and prepared reagents is stated in the corresponding chapters.

The strips of microtiter plate is stable up to the expiry date of the kit in the broken, but tightly closed bag when stored at 2–8°C.

Specimen Collection and Storage

Serum, Plasma (EDTA, Heparin)*

The usual precautions for venipuncture should be observed. It is important to preserve the chemical integrity of a blood specimen from the moment it is collected until it is assayed. Do not use grossly hemolytic icteric or grossly lipemic specimens. Samples appearing turbid should be centrifuged before testing to remove any particulate material.

Storage:	2-8°C	-20°C	Keep away from heat or direct sun light
Stability:	7 d	6 mon	Avoid repeated freeze-thaw cycles

Materials Supplied

1 x 12 x 8	MTP	Microtiter Plate Break apart strips. Microtiter plate with 12 rows each of 8 wells precoated with rituximab.
1 x 0.25 mL	RCTV CNTR	Reactive Control Ready-to-use. Contains rituximab-reactive reagent, human serum, stabilizers and <0.1% NaN ₃ .
1 x 0.5 mL	NEG CNTR	Negative Control Ready-to-use. Contains human serum, stabilizers and <0.1% NaN ₃ .
1 x 12 mL	ASSAY BUF	Assay Buffer Blue colored. Ready to use. Contains proteins and <0.1% NaN ₃ .
1 x 12 mL	POD CONJ	Peroxidase Conjugate Red colored. Ready to use. Contains peroxidase (POD) conjugated rituximab, stabilizer and preservatives.
1 x 12 mL	TMB SUBS	TMB Substrate Solution Ready to use. Contains TMB
1 x 12 mL	TMB STOP	TMB Stop Solution Ready to use. 1N HCl.
1 x 50 mL	WASHBUF CONC	Wash Buffer, Concentrate (20x) Contains Buffer with Tween 20.
2 x 1	ADH FILM	Adhesive Film For covering of Microtiter Plate during incubation

Materials Required But Not Supplied

1. Micropipettes (< 3% CV) and tips to deliver 5-1000 μ L.
2. Calibrated measures.
3. Tubes (1 mL) for sample dilution.
4. Wash bottle, automated or semi-automated microtiter plate washing system.
5. Microtiter plate reader capable of reading absorbance at 450/650 nm.
6. Bidistilled or deionised water, paper towels, pipette tips and timer.

Procedure Notes

1. Any improper handling of samples or modification of the test procedure may influence the results. The indicated pipetting volumes, incubation times, temperatures and pretreatment steps have to be performed strictly according to the instructions. Use calibrated pipettes and devices only.
2. Once the test has been started, all steps should be completed without interruption. Make sure that required reagents, materials and devices are prepared ready at the appropriate time. Allow all reagents and specimens to reach room temperature (18-25 °C) and gently swirl each vial of liquid reagent and sample before use. Mix reagents without foaming.
3. Avoid contamination of reagents, pipettes and wells/tubes. Use new disposable plastic pipette tips for each reagent, standard or specimen. Do not interchange caps. Always cap not used vials. Do not reuse wells/tubes or reagents.
4. Use a pipetting scheme to verify an appropriate plate layout.
5. Incubation time affects results. All wells should be handled in the same order and time sequences. It is recommended to use an 8-channel Micropipette or for pipetting of solutions in all wells.
6. Microplate washing is important. Improperly washed wells will give erroneous results. It is recommended to use a multi-channel pipette or an automatic microplate

washing system. Do not allow the wells to dry between incubations. Do not scratch coated wells during rinsing and aspiration. Rinse and fill all reagents with care. While rinsing, check that all wells are filled precisely with Wash Buffer, and that there are no residues in the wells.

- Humidity affects the coated wells/tubes. Do not open the pouch until it reaches room temperature. Unused wells/tubes should be returned immediately to the resealed pouch including the desiccant.

Preparation of Component

Dilute/dissolve	Component	with	Diluent	Ratio	Remarks	Storage	Stability
10 mL	Wash Buffer*	Up to 200 mL	bidist. Water	1:20	Warm up at 37°C to dissolve crystals. Mix vigorously.	2-8 °C	1 w

*Prepare Wash Buffer before starting assay procedure.

Test Procedure

1	Pipette 100µl of Assay Buffer non-exceptionally into each of the wells to be used.
2	Pipette 10 µL of ready-to use Negative Control, Reactive Control, and Samples into the respective wells of microtiter plate. Wells A1: Negative Control B1: Negative Control C1: Reactive Control D1 and on: Sample (Serum/Plasma)
3	Cover the plate with adhesive film. Briefly mix contents by gently shaking the plate. Incubate 60 min at room temperature (18-25°C).
4	Remove adhesive film. Discard incubation solution. Wash plate 3 times each with 300 µL of diluted Wash Buffer . Remove excess solution by tapping the inverted plate on a paper towel.
5	Pipette 100 µL of ready-to use Peroxidase Conjugate into each well.
6	Cover the plate with adhesive film. Incubate 60 min at room temperature (18-25°C).
7	Remove adhesive film. Discard incubation solution. Wash plate 3 times each with 300 µL of diluted Wash Buffer . Remove excess solution by tapping the inverted plate on a paper towel.

8	Pipette 100 µL of TMB Substrate Solution into each well.
9	Incubate 20 min (without adhesive foil.) at room temperature (18-25°C) in the dark .
10	Stop the substrate reaction by adding 100 µL of Stop Solution into each well. Briefly mix contents by gently shaking the plate. Color changes from blue to yellow.
11	Measure optical density with a photometer at 450/650 nm within 30 min after pipetting of the Stop Solution.

Interpretation Of Results

For the run to be valid, the OD_{450/650 nm} of Positive **Control** should be ≥ 1.00 and the OD_{450/650 nm} of each Negative Control should be < 0.200 , if not, improper technique or reagent deterioration may be suspected and the run should be repeated.

The results are evaluated by a cut-off value which is estimated by multiplying the mean OD_{450nm} of the negative controls by 3.
I.e.;

If "Sample OD_{450/650} the mean OD_{450/650} of Negative Controls" is ≥ 3 , the sample is POSITIVE. If "Sample OD_{450/650} the mean OD_{450/650} of Negative Controls" is < 3 , the sample is NEGATIVE.

References

1. Datasheet of Rituxan prescribing information
2. Datasheet of Mabthera prescribing information
3. Thurlings RM, Teng O, Vos K, Gerlag DM, Aarden L, Stapel SO, JM van Laar, Tak PP, Wolbink GJ, Clinical response, pharmacokinetics, development of human anti-chimaeric antibodies, and synovial tissue response to rituximab treatment in patients with rheumatoid arthritis, *Ann Rheum Dis* 2010; 69: 409–412.
4. Schmidt E, Hennig K, Mengede C, Zillikens D, Kromminga A, Immunogenicity of rituximab in patients with severe pemphigus, *Clinical Immunology*, 2009; 132: 334–341.
5. Keystone E, Fleischmann R, Emery P, Furst DE, van Vollenhoven R, Bathon J, Dougados M, Baldassare A, Ferraccioli G, Chubick A, Udell J, Cravets MW, Agarwal S, Cooper S, and Magrini F, Safety and Efficacy of Additional Courses of Rituximab in Patients With Active Rheumatoid Arthritis, *ARTHRITIS & RHEUMATISM*, 2007; 56: 3896–3908.
6. Shoko Goto, Hiroaki Goto, Reo Tanoshima, Hiromi Kato, Hiroyuki Takahashi, Osamu Sekiguchi, Sumio Kai, Serum sickness with an elevated level of human anti-chimeric antibody following treatment with rituximab in a child with chronic immune thrombocytopenic purpura, *Int J Hematol*, 2009; 89: 305–309.

7. Tobinai K, Y. Kobayashi, M. Narabayashi, M. Ogura, Y. Kagami, Y. Morishima, T. Ohtsu, T. Igarashi, Y. Sasaki, T. Kinoshita, T. Murate, Feasibility and pharmacokinetic study of a chimeric anti-CD20 monoclonal antibody (IDEC-C2B8, rituximab) in relapsed B-cell lymphoma, *Annals of Oncology* 1998; 9: 527-534.

8. Magda Nakou, Georgios Katsikas, Prodromos Sidiropoulos, George Bertias,

Eva Papadimitraki, Amalia Raptopoulou, Helen Koutala, Helen A Papadaki, Herakles Kriticos and Dimitrios T Boumpas, Rituximab therapy reduces activated B cells in both the peripheral blood and bone marrow of patients with rheumatoid arthritis: depletion of memory B cells correlates with clinical response *Arthritis Research & Therapy* 2009, 11:R131 (doi:10.1186/ar2798)

9. Pina M. Cardarelli, Maire Quinn, Dana Buckman, Yu Fang, David Colcher, David J. King, Christopher Bebbington, Geoffrey Yarranton, Binding to CD20 by Anti-B1 Antibody or F(ab')₂ is sufficient for induction of apoptosis in B-cell lines, *Cancer Immunology Immunotherapy*, 2002; 51: 15–24.

10. Maria Rehnberg, Sylvie Amu, Andrej Tarkowski, Maria I Bokarewa and Mikael Brisslert, Short- and long-term effects of anti-CD20 treatment on B cell ontogeny in bone marrow of patients with rheumatoid arthritis, *Arthritis Research & Therapy* 2009, 11:R123 (doi:10.1186/ar2789).

11. Ferdinand Breedveld, Sunil Agarwal, Ming Yin, Song Ren, Nicole F. Li, Tim M. Shaw, and Brian E. Davies, Rituximab Pharmacokinetics in Patients With Rheumatoid Arthritis: B-Cell Levels Do Not Correlate With Clinical Response, *Journal of Clinical Pharmacology*, 2007; 47: 1119-1128.

12. David G. Maloney, Antonio J. Grillo-López, Christine A. White, David Bodkin, Russell J. Schilder, James A. Neidhart, Nalini Janakiraman, Kenneth A. Foon, Tina-Marie Liles, Brian K. Dallaire, Ken Wey, Ivor Royston, Thomas Davis and Ronald Levy, IDEC-C2B8 (Rituximab) Anti CD20 Monoclonal Antibody Therapy in Patients With Relapsed Low-Grade Non-Hodgkin's Lymphoma, *Blood*, 1997; 90: 2188-2195.

13. Dieter Huhn, Christoph von Schilling, Martin Wilhelm, Anthony D. Ho, Michael Hallek, Rolf Kuse, Wolfgang Knauf, Ute Riedel, Axel Hinke, Stefanie Srock, Stefan Serke, Christian Peschel and Bertold Emmerich, Rituximab therapy of patients with B-cell chronic lymphocytic leukemia, *Blood*, 2001; 98: 1326-1331.

14. Maeda T, Yamada Y, Tawara M, Yamasaki R, Yakata Y, Tsutsumi C, Onimaru Y, Kamihira S, Tomonaga M., Successful treatment with a chimeric anti-CD20 monoclonal antibody (IDEC-C2B8, rituximab) for a patient with relapsed mantle cell lymphoma who developed a human anti-chimeric antibody. *Int J Hematol*. 2001; 74(1): 70-75.