



innovation for health & wellness

“trace & catch”

## Instruction for Use

Antibody to Vedolizumab (Entyvio®)

# SHIKARI® S-ATV

Enzyme immunoassay for the qualitative determination (screening) of antibodies to vedolizumab in serum and plasma

REF TR-ATVv1



12 x 8



2-8 C

Revision # 1.1 October 2018



Matriks Biotek Laboratories  
[www.matriksbiotek.com](http://www.matriksbiotek.com)

<b>Contents</b>	<b>Page</b>
Intend to use.....	3
Summary and Explanation.....	3
Test Principle .....	3
Warnings and Precautions .....	4
Storage and Stability.....	5
Specimen Collection and Storage.....	5
Materials Supplied.....	6
Materials Required but not Supplied .....	7
Procedure Notes.....	7
Preparation of Component .....	8
Test Procedure.....	9
Interpretation of Results.....	10
References.....	11

	SHIKARI Q-ATV
	Vedolizumab (Entyvio®) antibodies qualitative analyse
Required Volume (µl)	10
Total Time (min)	140
Sample	Serum, plazma
Sample Number	96
Detection Limit (ng/mL)	+/-
Spike Recovery (%)	-
Shelf Life (year)	1

## Intended Use

The Matriks Biotek Antibody to vedolizumab (Entyvio®) Enzyme-Linked-ImmunoSorbent-Assay (ELISA) Kit is intended for the qualitative determination of antibodies to vedolizumab (Entyvio®) in serum and plasma. It is for professional use only.

## Summary and Explanation

Vedolizumab is a recombinant humanized IgG1 monoclonal antibody directed against the human lymphocyte  $\alpha 4\beta 7$  integrin, a key mediator of gastrointestinal inflammation. It is used in the treatment of moderate to severe active ulcerative colitis and Crohn's disease for patients who have had an inadequate response with, lost response to, or were intolerant to inhibitors of tumor necrosis factor-alpha (TNF-alpha) or other conventional therapies. By blocking its primary target,  $\alpha 4\beta 7$  integrin, vedolizumab reduces inflammation in the gut.

Vedolizumab binds specifically to  $\alpha 4\beta 7$  integrin but does not bind to, or inhibit function of,  $\alpha 4\beta 1$  or  $\alpha E\beta 7$  integrins. Inhibition of the  $\alpha 4\beta 7$  integrin is a shared mechanism with natalizumab, however vedolizumab binds solely to the  $\alpha 4\beta 7$  but not the  $\alpha 4\beta 1$  integrin, unlike natalizumab which binds to both. As a result, natalizumab modulates the systemic immune system and is associated with other side effects such as progressive multifocal leukoencephalopathy (PML).

## Test Principle

The Matriks Biotek Antibody to infliximab (Remicade®) ELISA is a sandwich assay for the determination of antibodies against infliximab in serum and plasma samples. During the first incubation period, antibodies to infliximab (ATI) in patient serum/ plasma samples are captured by the drug infliximab (Remicade®) coated on the wall of the microtiter wells. After washing away the unbound components from samples, a peroxidase-labelled specific conjugate is added to each well and then incubated.

After a second washing step, the bound enzymatic activity is detected by addition of tetramethylbenzidine (TMB) chromogen-substrate. Finally, the reaction is terminated with an acidic stop solution. The intensity of the reaction color is directly proportional to the concentration of ATI in sample.

## Warnings and Precautions

1. For professional use only.
2. Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood. For further information (clinical background, test performance, automation protocols, alternative applications, literature, etc.) please refer to the local distributor.
3. In case of severe damage of the kit package please contact Matriksbiotek or your supplier in written form, latest one week after receiving the kit. Do not use damaged components in test runs, but keep safe for complaint related issues.
4. Obey lot number and expiry date. Do not mix reagents of different lots. Do not use expired reagents.
5. Follow good laboratory practice and safety guidelines. Wear lab coats, disposable latex gloves and protective glasses where necessary.
6. Reagents of this kit containing hazardous material may cause eye and skin irritations. See MATERIALS SUPPLIED and labels for details.
7. Chemicals and prepared or used reagents have to be treated as hazardous waste according the national biohazard safety guidelines or regulations.
8. Avoid contact with Stop solution. It may cause skin irritations and burns.
9. Some reagents contain sodium azide ( $\text{NaN}_3$ ) as preservatives. In case of contact with eyes or skin, flush immediately with water.  $\text{NaN}_3$  may react with lead and copper plumbing to form explosive metal azides. When disposing reagents, flush with large volume of water to avoid azide build-up.
10. All reagents of this test kit containing human serum or plasma have been tested and were found negative for HIV I/II, HBsAg and HCV by FDA approved procedures. However, a presence of these or other infectious agents cannot be excluded absolutely and therefore reagents should be treated as potential biohazards in use and for disposal.

## Storage and Stability

The kit is shipped at ambient temperature and should be stored at 2-8°C. Keep away from heat or direct sun light. The storage and stability of specimen and prepared reagents is stated in the corresponding chapters. The strips of microtiter plate is stable up to the expiry date of the kit in the broken, but tightly closed bag when stored at 2-8°C.

## Specimen Collection and Storage

### Serum, Plasma (EDTA, Heparin)\*

The usual precautions for venipuncture should be observed. It is important to preserve the chemical integrity of a blood specimen from the moment it is collected until it is assayed. Do not use grossly hemolytic, icteric or grossly lipemic specimens. Samples appearing turbid should be centrifuged before testing to remove any particulate material.

Storage:	2-8°C	-20°C	Keep away from heat or direct sun light
Stability:	7 d	6 mon	Avoid repeated freeze-thaw cycles

\*. Vedolizumab (Entyvio®) infusion camouflages/masks the presence of antibody to vedolizumab (ATV) in serum/plasma samples. Therefore, blood sampling time is critical for detection of ATV. Matriks Biotek Laboratories propose to obtain blood sample just before the infusion of vedolizumab (Entyvio®) or at least 2 weeks after the infusion of infliximab (Remicade®).

## Materials Supplied

1 x 12 x 8	MTP	<b>Microtiter Plate</b> Break apart strips. Microtiter plate with 12 rows each of 8 wells coated with vedolizumab.
1 x 0.25 mL	RCTV CNTR	<b>Reactive Control</b> Ready-to-use. Contains vedolizumab-reactive reagent, human serum, stabilizers and <0.1% NaN <sub>3</sub>
1 x 0.5 mL	NEG CNTR	<b>Negative Control</b> Ready-to-use. Contains human serum, stabilizers and <0.1% NaN <sub>3</sub>
1 x 12 mL	ASSAY BUF	<b>Assay Buffer</b> Blue colored. Ready to use. Contains proteins and <0.1% NaN <sub>3</sub>
1 x 12 mL	POD CONJ	<b>Peroxidase Conjugate</b> Red colored. Ready to use. Contains peroxidase (POD) conjugate, stabilizer and preservatives.
1 x 12 mL	TMB SUBS	<b>TMB Substrate Solution</b> Ready to use. Contains TMB
1 x 12 mL	TMB STOP	<b>TMB Stop Solution</b> Ready to use. 1N HCl.
1 x 50 mL	WASHBUF CONC	<b>Wash Buffer, Concentrate (20x)</b> Contains Buffer with Tween 20.
2 x 1	ADH FILM	<b>Adhesive Film</b> For covering of Microtiter Plate during incubation.

## Materials Required but not Supplied

1. Micropipettes (< 3% CV) and tips to deliver 5-1000  $\mu\text{L}$ .
2. Calibrated measures.
3. Tubes (1 mL) for sample dilution.
4. Wash bottle, automated or semi-automated microtiter plate washing system.
5. Microtiter plate reader capable of reading absorbance at 450/650 nm.
6. Bidistilled or deionised water, paper towels, pipette tips and timer.

## Procedure Notes

1. Any improper handling of samples or modification of the test procedure may influence the results. The indicated pipetting volumes, incubation times, temperatures and pretreatment steps have to be performed strictly according to the instructions. Use calibrated pipettes and devices only.
2. Once the test has been started, all steps should be completed without interruption. Make sure that required reagents, materials and devices are prepared ready at the appropriate time. Allow all reagents and specimens to reach room temperature (18-25 °C) and gently swirl each vial of liquid reagent and sample before use. Mix reagents without foaming.
3. Avoid contamination of reagents, pipettes and wells/tubes. Use new disposable plastic pipette tips for each reagent, standard or specimen. Do not interchange caps. Always cap not used vials. Do not reuse wells/tubes or reagents.
4. Use a pipetting scheme to verify an appropriate plate layout.
5. Incubation time affects results. All wells should be handled in the same order and time sequences. It is recommended to use an 8-channel Micropipettor for pipetting of solutions in all wells.
6. Microplate washing is important. Improperly washed wells will give erroneous results. It is recommended to use a multichannel pipette or an automatic microplate washing system. Do not allow the wells to dry between incubations. Do not scratch coated wells during rinsing and aspiration. Rinse and fill all reagents with care. While rinsing, check that all wells are filled precisely with Wash Buffer, and that there are no residues in the wells.
7. Humidity affects the coated wells/tubes. Do not open the pouch until it reaches room temperature. Unused wells/tubes should be returned immediately to the resealed pouch including the desiccant.

## Preparation of Component

Dilute/ dissolve	Component	with	Diluent	Relation	Remarks	Storage	Stability
10 mL	Wash Buffer*	Up to 200 mL	Bidist. water	1:20	Warm up at 37°C to dissolve crystals. Mix vigorously.	2-8 °C	2 w

\*. Prepare Wash Buffer before starting assay procedure.



## Test Procedure

1	Pipette 100µl of Assay Buffer non-exceptionally into each of the wells to be used.
2	<p><b>QUALITATIVE ELISA TEST FORMAT</b></p> <p>Pipette 10 µL of ready-to use Negative Control, Reactive Control, and Samples into the respective wells of microtiter plate.</p> <p><u>Wells</u></p> <p>A1:                      Negative Control            B1:                      Negative Control            C1:                      Reactive Control            D1 and on.:            Sample (Serum/Plasma)</p>
3	Cover the plate with adhesive film. Briefly mix contents by gently shaking the plate. Incubate 60 min at room temperature (18-25°C).
4	Remove adhesive film. Discard incubation solution. Wash plate 3 times each with 300 µL of diluted Wash Buffer. Remove excess solution by tapping the inverted plate on a paper towel.
5	Pipette 100 µL of ready-to use Peroxidase Conjugate into each well.
6	Cover the plate with adhesive film. Incubate 60 min at room temperature (18- 25°C).
7	Remove adhesive film. Discard incubation solution. Wash plate 3 times each with 300 µL of diluted Wash Buffer. Remove excess solution by tapping the inverted plate on a paper towel.
8	Pipette 100 µL of TMB Substrate Solution into each well.
9	Incubate 20 min (without adhesive foil.) at room temperature (18-25°C) in the dark.
10	Stop the substrate reaction by adding 100 µL of Stop Solution into each well. Briefly mix contents by gently shaking the plate. Color changes from blue to yellow
11	Measure optical density with a photometer at 450/650 nm within 30 min after pipetting of the Stop Solution.

## Interpretation Of Results

For the run to be valid, the OD<sub>450/650</sub> nm of Positive Control should be  $\geq 1.00$  and the OD<sub>450/650</sub> nm of each Negative Control should be  $< 0.2000$ , if not, improper technique or reagent deterioration may be suspected and the run should be repeated.

**The results are evaluated by a cut-off value which is estimated by multiplying the mean OD<sub>450/650</sub> nm of the negative controls by 3.**

E.g.;

**If "Sample OD<sub>450/650</sub> the mean OD<sub>450/650</sub> of Negative Controls" is  $\geq 3$ , the sample is POSITIVE**

**If "Sample OD<sub>450/650</sub> the mean OD<sub>450/650</sub> of Negative Controls" is  $< 3$ , the sample is NEGATIVE**

## References

1. Ardizzone S, Bianchi Porro G: Biologic therapy for inflammatory bowel disease. *Drugs*. 2005;65(16):2253-86. [PubMed:16266194]
2. Soler D, Chapman T, Yang LL, Wyant T, Egan R, Fedyk ER: The binding specificity and selective antagonism of vedolizumab, an anti-alpha4beta7 integrin therapeutic antibody in development for inflammatory bowel diseases. *J Pharmacol Exp Ther*. 2009 Sep;330(3):864-75. doi: 10.1124/jpet.109.153973. Epub 2009 Jun 9. [PubMed:19509315]
3. Krupka N, Baumgart DC: Designing biologic selectivity for inflammatory bowel disease--role of vedolizumab. *Drug Des Devel Ther*. 2014 Dec 17;9:147-54. doi: 10.2147/DDDT.S50348. eCollection 2015. [PubMed:25552903]
4. Wang MC, Zhang LY, Han W, Shao Y, Chen M, Ni R, Wang GN, Wei FX, Zhang YW, Xu XD, Zhang YC: PRISMA--efficacy and safety of vedolizumab for inflammatory bowel diseases: a systematic review and meta-analysis of randomized controlled trials. *Medicine (Baltimore)*. 2014 Dec;93(28):e326. doi: 10.1097/MD.0000000000000326. [PubMed:25526490]
5. Maria Rosario, Nathanael L. Dirks, Catherine Milch, Asit Parikh, Michael Bargfrede, Tim Wyant, Eric Fedyk, Irving Fox A Review of the Clinical Pharmacokinetics, Pharmacodynamics, and Immunogenicity of Vedolizumab *Clin Pharmacokinet*. 2017; 56(11): 1287–1301. Published online 2017 May 18. doi: 10.1007/s40262-017-0546-0